

PCTWORLD INTELLECTUAL PROPERTY ORGANIZATION
International Bureau

INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

(51) International Patent Classification ⁷ : C12Q 1/00	A2	(11) International Publication Number: WO 00/55354 (43) International Publication Date: 21 September 2000 (21.09.00)
<p>(21) International Application Number: PCT/GB00/01005</p> <p>(22) International Filing Date: 17 March 2000 (17.03.00)</p> <p>(30) Priority Data: 9906140.0 17 March 1999 (17.03.99) GB</p> <p>(71) Applicant (for all designated States except US): UNIVERSITY OF BRISTOL [GB/GB]; Senate House, Tyndall Avenue, Bristol BS8 1TH (GB).</p> <p>(72) Inventors; and (75) Inventors/Applicants (for US only): SOOTHILL, Peter [GB/GB]; University of Bristol, Fetal Medicine Research Unit, St. Michaels Hill, Southwell Street, Bristol BS2 8ES (GB). MILLAR, Michael [GB/GB]; University of Bristol, Senate House, Tyndall Avenue, Bristol BS8 1TH (GB). CORFIELD, Tony [GB/GB]; University of Bristol, Dorothy Crowfoot Hodgkin Labs., Division of Medicine, Bristol Royal Infirmary, Marlborough Street, Bristol BS2 8HU (GB).</p> <p>(74) Agent: NASH, David, Allan; Haseltine Lake & Co., Imperial House, 15-19 Kingsway, London WC2B 6UD (GB).</p>		<p>(81) Designated States: AE, AG, AL, AM, AT, AU, AZ, BA, BB, BG, BR, BY, CA, CH, CN, CR, CU, CZ, DE, DK, DM, DZ, EE, ES, FI, GB, GD, GE, GH, GM, HR, HU, ID, IL, IN, IS, JP, KE, KG, KP, KR, KZ, LC, LK, LR, LS, LT, LU, LV, MA, MD, MG, MK, MN, MW, MX, NO, NZ, PL, PT, RO, RU, SD, SE, SG, SI, SK, SL, TJ, TM, TR, TT, TZ, UA, UG, US, UZ, VN, YU, ZA, ZW, ARIPO patent (GH, GM, KE, LS, MW, SD, SL, SZ, TZ, UG, ZW), Eurasian patent (AM, AZ, BY, KG, KZ, MD, RU, TJ, TM), European patent (AT, BE, CH, CY, DE, DK, ES, FI, FR, GB, GR, IE, IT, LU, MC, NL, PT, SE), OAPI patent (BF, BJ, CF, CG, CI, CM, GA, GN, GW, ML, MR, NE, SN, TD, TG).</p> <p>Published <i>Without international search report and to be republished upon receipt of that report.</i></p>

(54) Title: A DIAGNOSTIC TEST

(57) Abstract

There is provided a method for the detection of a disease comprising the following steps: (i) obtaining a biological sample from a human subject; (ii) applying the biological sample to a sialidase substrate which has been immobilised on a solid support medium; and (iii) detecting a change in the immobilised sialidase substrate. The substrate is preferably 5-bromo-4-chloro-3-indoyl α -D-N-acetyl neuraminic acid or a salt thereof.

BEST AVAILABLE COPY

A DIAGNOSTIC TEST

This invention relates to a method for the detection of a disease, a device for use in such a method and a kit comprising such a device.

5 Bacterial vaginosis (BV) is a condition in which the most numerous micro-inhabitants of the vagina (*Lactobacilli*) are overwhelmed by other micro-organisms such as *Gardnerella vaginalis*. BV is associated with preterm labour (PTL). The mechanism behind this
10 association is thought to be due to BV-related flora gaining access to the uterus and causing the release of inflammatory mediators that initiate labour.

It has previously been thought that in BV, the bacterial infection itself gives rise to PTL.
15 Antibiotics are commonly used to treat BV, for example to prevent PTL, but such treatment is rarely successful. Moreover antibiotics can remove non-pathogenic bacteria, including beneficial strains such as *Lactobacilli*.

20 At present, the presence of BV is detected on the basis of a clinical examination and Gram staining for the bacteria characteristic of the disease. It takes a few days in order to get a result using Gram staining so the patient cannot be told the result at the time of
25 visiting the clinic.

During pregnancy, there is no membranous barrier between the vagina and the uterus to prevent an ascending bacterial infection. However, the cervical mucus plug is thought to act as a physical barrier to
30 in-utero infection.

The mucus plug is built on an infrastructure of carbohydrate-rich glycoproteins called mucins. Mucins are composed of oligosaccharide side chains joined to a central protein core. Sialic acid residues are present
35 at the terminal ends of the carbohydrate side-chains.

High levels of bacterial sialidase activity are found in the vaginal secretions of women with BV as opposed to those without BV (MacGregor et al (1994) Am. J. Obstet. Gynaecol. 170 1048-1060; Briselden et al (1992) J. Clin. Microbiol. 30 663-666).

Bacterial sialidases are enzymes which cleave sialic acid residues of glycolipids and glycoproteins. To date, sialidase activity has been assayed by colorimetric, fluorimetric and radioactive techniques using a variety of natural and synthetic sialoglycoconjugate substrates. Substrates which have previously been used in assays for sialidase activity include: bovine submandibular gland mucin and human α -1 acid glycoprotein (Howe et al (1999) in press); 2'-(4-methylumbelliferyl)- α -D-N-acetylneuraminic acid (Paupermpoonsiri et al (1996) Clin. Inf. Dis. Vol. 23 748-752; Brieselden et al (1992) *supra*); 4-nitrophenyl- α -sialic acid and 2-(3-methoxyphenyl)-N-acetyl-D-neuraminic acid (MacGregor et al (1994) *supra*). The sensitivity of many of the known sialidase substrates is low.

The sialidase substrates described to date are used in a variety of assays. For example, colorimetric or fluorimetric assays may be used with synthetic substrates, which detect the non-sialic moiety of the substrate (for example 4-methyl umbelliferone or 4-nitrophenol). Alternatively, the amount of sialic acid released from a sialoglycoconjugate substrate can be detected by HPLC or a colorimetric assay. Also, radioactive assays are sometimes used which involve the introduction of a radiolabel into the sialoconjugate substrate and detection of the released sialic acid by liquid scintillation counting.

The present inventors have discovered that the link between PTL and BV is due to bacterial sialidase

activity rather than bacterial infection itself. It is believed that sialidase activity is involved in the degradation of the mucus plug, which results in increased bacterial access into the upper reproductive tract.

The carboxyl groups of the sialic acids at the ends of the carbohydrate side-chains of the mucin molecules confer a negative ionic charge causing rigidity of the sugar side chains. Cleavage of sialic acid destroys the mutual repulsive charge between the mucin molecules causing a loss of viscosity of the mucus plug. Bacterial access into the upper reproductive tract is increased as a result of this action for two reasons. Firstly, mucin and therefore mucus matrix organisation is lost after degradation and the mechanical and bacteriostatic properties of the mucus matrix are rendered less effective as barrier mechanisms. Secondly, disintegration of the mucus layer facilitates bacterial adherence to the underlying epithelium which alters immunological recognition responses and reduces the likelihood that bacteria will be washed out by the movement of the vaginal fluid.

It follows from this discovery that sialidase inhibitors should be useful to treat and prevent conditions such as bacterial vaginosis, and will be useful in therapies to prevent PTL. It also follows that assays for sialidase activity will be useful to detect diseases such as BV.

The present inventors have also developed a diagnostic test for diseases such as BV which detects the presence of sialidase activity. A result in this test has been found to correlate significantly with the incidence of BV using a conventional Gram-stain diagnosis. The test method involves taking a sample from a patient and contacting the sample with a sialidase substrate. After incubation, detectable

change such as a colour change, is observed if sialidase activity is detected. In a preferred embodiment, the test is a spot test, that is to say that the sample is applied to the sialidase substrate which is itself supported by a suitable solid medium, such as a filter paper.

The test makes use of substrates which have been found to have a very high sensitivity in the detection of BV, indicating possibly an optimum environment for the binding of sialidases from BV flora. The preferred substrate used in the present test is a salt of 5-bromo-4-chloro-3-indoyl α -D-N-acetyl neuraminic acid, in particular the cyclohexyl amine (X- α -NANA), which is available for Rose Scientific Ltd., Edmonton, Canada.

The test method of the present invention for detecting sialidase activity is highly sensitive, inexpensive to run and quick and simple to use. The test is especially useful to test for sialidase activity as an indicator of BV and therefore a predictor of the likelihood of preterm birth. The assay can be carried out in the same room as the patient or in a clinic or GP surgery, since the method of the invention does not require the use of items of expensive equipment or materials other than those which would normally be available. For example, in contrast to conventional methods, there is no need for a Gram staining kit or a light microscope. Moreover, the method of the invention is less labour intensive than a Gram stain.

Also, the sialidase spot test has an advantage over the colorimetric, fluorimetric or radioactivetests which are currently used to detect BV in that, since sialidase activity is not always present in BV or in all BV flora, its presence may be an indicator that the condition is more detrimental to the host in some cases than in others.

According to a first aspect of the invention there is provided a method for the detection of a disease comprising the following steps:

- 5 (i) obtaining a biological sample from a human subject;
- (ii) applying the biological sample to a sialidase substrate which has been immobilised on a solid support medium; and
- 10 (iii) detecting a change in the immobilised sialidase substrate.

According to a second aspect there is provided a device for use in the detection of a disease, which device comprises a sialidase-specific substrate immobilised on a solid medium.

15 The device may be used in a method according to the first aspect of the invention, which comprises the following steps:

- i) obtaining a biological sample from a human subject;
- 20 ii) applying the biological sample to the device; and
- iii) detecting the presence or absence of a colour change in the immobilised sialidase substrate

According to a third aspect of the invention there is provided a kit for the diagnosis of a disease which comprises a device according to the second aspect of the invention.

25 The kit of this third aspect of the invention may also comprise an incubation medium, for example a Tris/HCl buffer and/or a sample-obtaining means.

30 Preferably the device of the second aspect of the invention and the kit of the third aspect of the invention is used to diagnose a genito-urinary tract infection. More preferably the disease is bacterial
35 vaginosis.

The sialidase substrate is a substrate which is

recognised specifically by a sialidase enzyme. For example, the substrate may be cleaved by the sialidase enzyme. Preferably the substrate is a glycoside of sialic acid. More preferably the substrate is 5-bromo-4-chloro-3-indoyl α -D-N-acetyl neuraminic acid or a salt thereof, for example the cyclohexyl amine salt thereof.

The solid medium may be, for example, filter paper.

Preferably the mammalian subject is a pregnant woman or a woman who is contemplating pregnancy.

According to a fourth aspect of the present invention, there is provided a method for the detection of sialidase activity in a sample, which method comprises the following steps:

- (a) contacting said sample with a test reagent comprising 5-bromo-4-chloro-3-indoyl α -D-N-acetyl neuraminic acid or a salt thereof; and
- (b) detecting a change in the test reagent.

The test reagent may comprise, for example, a solid, liquid or gel phase medium on or in which the 5-bromo-4-chloro-3-indoyl α -D-N-acetyl neuraminic acid or salt thereof is supported, applied, adsorbed or otherwise contained.

The 5-bromo-4-chloro-3-indoyl α -D-N-acetyl neuraminic acid is preferably in the form of the cyclohexylamine salt. Preferably, the 5-bromo-4-chloro-3-indoyl α -D-N-acetyl neuraminic acid cyclohexyl amine is supported or adsorbed on a solid substrate, for example a filter paper. In this way, the reactive compound may be immobilised on the substrate to facilitate the carrying out of the diagnostic test. Such a diagnostic test where the reactive compound is adsorbed on a solid medium is often referred to as a spot-test, since the diagnostic test itself is carried out by "spotting" the sample (or an extract of the

sample) onto the solid medium in the area where the reactive compound has been adsorbed.

The sample may be obtained directly from a mammalian subject, typically a human subject, for example as a swab, or a washing using a suitable buffer. After the sample has been taken, it may be applied to or contacted directly with the test reagent. Alternatively, for example where the sample is a swab, the contents of the sample may be extracted in a liquid medium, for example a buffer such as tris buffer and then applied to the test reagent. This may make application of the test method more convenient.

The test method is particularly suited for the detection of BV, and so, in a preferred embodiment, the sample is taken as a vaginal swab or sample from a female human patient.

After the sample has been contacted with the test reagent, any change in the test reagent is observed. In the case where the test reagent comprises 5-bromo-4-chloro-3-indoyl α -D-N-acetyl neuraminic acid cyclohexyl amine, the change is a colour change, from colourless to blue. This change may be observed visually, or may be carried out using suitable equipment, such as a stick test colorimeter or a spectrophotometer, which could be used to quantify the result. Where a colour change is to be observed, it is preferred that the background part of the substrate is not of a colour which would obscure or mask the colour change in the reactive compound. It is also preferred that the unreacted test reagent is of a colour which does not show up on the background part of the substrate. Preferably the unreacted test reagent is colourless.

According to a fifth aspect there is provided a kit for use in the fourth aspect of the present invention, which comprises a test reagent comprising 5-bromo-4-chloro-3-indoyl α -D-N-acetyl neuraminic acid or

a salt thereof.

5 In this embodiment of the invention, the test reagent is preferably a device in the form of a solid medium on which the 5-bromo-4-chloro-3-indoyl α -D-N-acetyl neuraminic acid or salt thereof is supported or adsorbed. The preferred test reagent is a filter paper soaked in 5-bromo-4-chloro-3-indoyl α -D-N-acetyl neuraminic acid cyclohexyl amine solution. Preferably the filter paper is soaked in a solution of 5-bromo-4-chloro-3-indoyl α -D-N-acetyl neuraminic acid cyclohexyl amine at approximately 1mg/ml. The filter paper can then be used directly or dried and stored for future use. Hence the kit can either comprise 5-bromo-4-chloro-3-indoyl α -D-N-acetyl neuraminic acid cyclohexyl amine (for example in solution) and a solid medium (for example filter paper) such that the user can apply the substrate to the solid medium, or the kit can comprise a solid medium to which 5-bromo-4-chloro-3-indoyl α -D-N-acetyl neuraminic acid cyclohexyl amine solution has been previously applied and dried.

15 The kit of this aspect of the invention may also comprise an incubation medium, for example a Tris/HCl buffer and/or a sample-obtaining means, and may be packaged together in a suitable container with appropriate instructions for use.

20 The test reagent used in the present invention may be prepared by a process in which the reactive compound is brought into contact with a suitable solid, liquid or gel phase medium so that the reactive compound is contained in or supported or adsorbed on the medium. For example, where the test reagent is a solid device, the reactive compound may be prepared in a suitable solution which is then applied to the solid medium.

25 The test reagent employed in the fourth and fifth aspects of the present invention is itself an aspect of the present invention. A preferred form of the reagent

is a solid device in which the test reagent is applied to a surface of a solid substrate to form a local patch which is where the diagnostic test is carried out.

5 The method of the fourth aspect of the invention, the kit of the fifth aspect of the invention and the device of the sixth aspect of the invention may be used to detect a disease which is characterised by abnormal (for example increased) activity of bacterial
10 sialidases. For example, the disease may be a genito-uninary tract infection. Preferably the method of the fourth aspect of the invention, the device of the fifth aspect of the invention and the kit of the sixth aspect of the invention is used to diagnose bacterial
15 vaginosis.

Preferably the mammalian subject is a pregnant woman or a woman who is contemplating pregnancy.

According to a seventh aspect of the invention there is provided a method for diagnosing PTL in a human subject which method comprises the step of
20 assaying bacterial sialidase activity in the patient or in a sample taken from the patient.

According to a eighth aspect of the invention there is provided a method for treating BV or preventing PTL which method comprises the step of administering a
25 sialidase inhibitor to the patient.

According to a ninth aspect of the invention there is provided the use of a sialidase inhibitor in the manufacture of a therapeutic or prophylactic composition for the treatment of BV or for the
30 prevention of PTL.

For a better understanding of the present invention, and to show how the same may be put into effect, reference will now be made, by way of example only, to the following example.

35 **EXAMPLE**

A parallel Gram-stain and sialidase spot-test assay for

the diagnosis of BV

100 women attending the Milne Genito-Urinary Clinic, Bristol Royal Infirmary, were tested for sialidase activity in addition to the routine Gram stain for the detection of BV. A separate swab was taken at the time of examination and placed immediately into a buffer composed of 2mls 25mM Tris/HCl/Tween 20 (pH 7.0). The swabs were placed in cold store at 4°C until the end of clinic.

The sialic acid substrate 5-Bromo-4-chloro-3-indolyl- α -D-N-acetylneuraminic acid cyclohexylamine salt was weighed out to a concentration of 1mg/ml in a buffer of 25mM CaCl₂, 150mM Na acetate and 1M NaCl (pH 5.5). This was mixed well and then pipetted onto a 10cm diameter piece of filter paper until the filter paper was well damped. The swabs were removed from the Tris buffer and rubbed onto the filter paper inoculated with substrate. The filter paper was placed in a Petri dish, the lid closed and the dish incubated for 1 hour at 37°C. A blue spot would appear indicating sialidase activity was present. If there was no sialidase activity, the paper remained colourless even after a period of 2 hours prior to the filter paper drying out.

Gram stain diagnosis of BV was assessed by Spiegel's criteria (1980). Grade 2 BV as defined by Hay et al (1992) is a problematic category whose relationship to both BV and "normal" vaginal flora needs further elucidation. This category does not fulfil the definition of BV and it is unclear whether Grade 2 BV is an indication that colonization with BV microorganisms undergoes a stepwise progression. In this study, therefore, Grades 1 and 2 were pooled for the purposes of the statistical analysis, although initial assessment of the Gram stains divided them into Grade 1 (predominantly *Lactobacillus* morphotypes), Grade 2 (mixed *Lactobacillus* and organisms associated

with BV) and Grade 3 (few or no *Lactobacilli*, predominantly BV organisms with the presence of clue cells). Observers were blinded to the results of the spot test and spot test observers were blinded to the results of the Gram stain.

Results

Gram stain results for BV were correlated with the results of the spot test. The results of Grade 1 and Grade 2 were pooled for the statistical analysis, as Grade 2 is not regarded as a definitive diagnosis of BV and the numbers in this study were relatively small.

The patient's ages ranges were 16 years to 52 years with the mean age being 28.3 years, the median 26 years and the mode 25 years.

A positive spot test for sialidase activity was significantly correlated with the incidence of BV Grade 3 on Gram stain diagnosis at the $p < .000001$ level. The sensitivity of the spot test in the detection of BV was therefore 93.5 and the specificity 96.3. The positive predictive value was 95.6 and the negative predictive value 94.5. Amsells (1990) criteria was significantly associated with the results of the spot test at the $p = .04$ level, with the sensitivity being 93.7, specificity 70.5, the positive predictive value 57.6 and the negative value 96.2.

CLAIMS

1. A method for the detection of a disease comprising the following steps:

5 (i) obtaining a biological sample from a human subject;

(ii) applying the biological sample to a sialidase substrate which has been immobilised on a solid support medium; and

10 (iii) detecting a change in the immobilised sialidase substrate.

2. A method according to claim 1, wherein the disease is bacterial vaginosis.

3. A method according to claim 1, wherein the substrate is 5-bromo-4-chloro-3-indoyl α -D-N-acetyl neuraminic acid or a salt thereof.

15 4. A device for use in method for detecting a disease according to any preceding claim.

5. A kit for the diagnosis of a disease which comprises the device according to claim 4.

20 6. A method for the detection of sialidase activity in a sample, which method comprises the following steps:

25 (a) contacting said sample with a test reagent comprising 5-bromo-4-chloro-3-indoyl α -D-N-acetyl neuraminic acid or a salt thereof; and

(b) detecting a change in the test reagent.

7. A method according to claim 6, wherein the 5-bromo-4-chloro-3-indoyl α -D-N-acetyl neuraminic acid is adsorbed on a solid medium.

30 8. A method according to claim 6 or 7, wherein the method is used to detect bacterial vaginosis.

INTERNATIONAL SEARCH REPORT

International Application No

PCT/GB 00/01005

A. CLASSIFICATION OF SUBJECT MATTER

IPC 7 C12Q1/34 C12Q1/00 G01N33/50

According to International Patent Classification (IPC) or to both national classification and IPC

B. FIELDS SEARCHED

Minimum documentation searched (classification system followed by classification symbols)

IPC 7 C12Q G01N

Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched

Electronic data base consulted during the international search (name of data base and, where practical, search terms used)

PAJ, EPO-Internal

C. DOCUMENTS CONSIDERED TO BE RELEVANT

Category *	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
X	CHEMICAL ABSTRACTS, vol. 110, no. 11, 13 March 1989 (1989-03-13) Columbus, Ohio, US; abstract no. 90993t, MONCLA B J ET AL: "Detection of sialidase (neuraminidase) activity in Actinomyces species by using 2'-(4-methylumbelliferyl)alpha-D-N-acetyl euraminic acid in a filter paper spot test" page 312; column 2; XP002901152 abstract & J CLIN MICROBIOL, vol. 27, no. 1, 1989, pages 182-184, --- -/--	1,4,5

☒ Further documents are listed in the continuation of box C.☒ Patent family members are listed in annex.

* Special categories of cited documents :

- *A* document defining the general state of the art which is not considered to be of particular relevance
- *E* earlier document but published on or after the international filing date
- *L* document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)
- *O* document referring to an oral disclosure, use, exhibition or other means
- *P* document published prior to the international filing date but later than the priority date claimed

- *T* later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention
- *X* document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone
- *Y* document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art
- *Z* document member of the same patent family

Date of the actual completion of the international search

4 July 2000

Date of mailing of the international search report

31.08.2000

Name and mailing address of the ISA

European Patent Office, D-53115 Bonn, Germany

Authorized officer

INTERNATIONAL SEARCH REPORT

International Application No

PCT/GB 00/01005

C.(Continuation) DOCUMENTS CONSIDERED TO BE RELEVANT

Category *	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
X	CHEMICAL ABSTRACTS, vol. 120, no. 15, 11 April 1994 (1994-04-11) Columbus, Ohio, US; abstract no. 192216r, FUJII I: "Preparation of 2-O-(5-bromo-4-chloroindol-3-yl) sialic acid" page 1103; column 2; XP002901153 abstract & JP 05 339283 A (TANPAKU KOGAKU KENKYUSHO KK) 21 December 1993 (1993-12-21)	6
X	BRISLEDEN A M ET AL: "Sialidases (Neuraminidases) in Bacterial Vaginosis and Bacterial Vaginosis-Associated Microflora" JOURNAL OF CLINICAL MICROBIOLOGY, vol. 30, no. 3, March 1992 (1992-03), pages 663-666, XP002901154 abstract page 664, left-hand column, paragraph 3	1,2,4
A	MCGREGOR J A ET AL: "Bacterial vaginosis is associated with prematurity and vaginal fluid mucinase and sialidase: Results of a controlled trial of topical clindamycin cream" AMERICAN JOURNAL OF OBSTETRICS AND GYNECOLOGY, vol. 170, no. 4, April 1994 (1994-04), pages 1048-1060, XP002901155 abstract	1,2
A	US 4 355 102 A (QUASH GERARD) 19 October 1982 (1982-10-19) claim 1	1,4,5

INTERNATIONAL SEARCH REPORT

Information on patent family members

International Application No

PCT/GB 00/01005

Patent document cited in search report	Publication date	Patent family member(s)	Publication date
JP 05339283 A	21-12-1993	JP 2904647 B	14-06-1999
US 4355102 A	19-10-1982	FR 2450877 A	03-10-1980
		AT 1480 T	25-08-1982
		DE 3060771 D	21-10-1982
		EP 0015841 A	17-09-1980
		ES 489267 D	01-11-1981
		ES 8201209 A	01-03-1982
		JP 55141198 A	04-11-1980

FOR THE PURPOSES OF INFORMATION ONLY

Codes used to identify States party to the PCT on the front pages of pamphlets publishing international applications under the PCT.

AL	Albania	ES	Spain	LS	Lesotho	SI	Slovenia
AM	Armenia	FI	Finland	LT	Lithuania	SK	Slovakia
AT	Austria	FR	France	LU	Luxembourg	SN	Senegal
AU	Australia	GA	Gabon	LV	Latvia	SZ	Swaziland
AZ	Azerbaijan	GB	United Kingdom	MC	Monaco	TD	Chad
BA	Bosnia and Herzegovina	GE	Georgia	MD	Republic of Moldova	TG	Togo
BB	Barbados	GH	Ghana	MG	Madagascar	TJ	Tajikistan
BE	Belgium	GN	Guinea	MK	The former Yugoslav Republic of Macedonia	TM	Turkmenistan
BF	Burkina Faso	GR	Greece	ML	Mali	TR	Turkey
BG	Bulgaria	HU	Hungary	MN	Mongolia	TT	Trinidad and Tobago
BJ	Benin	IE	Ireland	MR	Mauritania	UA	Ukraine
BR	Brazil	IL	Israel	MW	Malawi	UG	Uganda
BY	Belarus	IS	Iceland	MX	Mexico	US	United States of America
CA	Canada	IT	Italy	NE	Niger	UZ	Uzbekistan
CF	Central African Republic	JP	Japan	NL	Netherlands	VN	Viet Nam
CG	Congo	KE	Kenya	NO	Norway	YU	Yugoslavia
CH	Switzerland	KG	Kyrgyzstan	NZ	New Zealand	ZW	Zimbabwe
CI	Côte d'Ivoire	KP	Democratic People's Republic of Korea	PL	Poland		
CM	Cameroon	KR	Republic of Korea	PT	Portugal		
CN	China	KZ	Kazakhstan	RO	Romania		
CU	Cuba	LC	Saint Lucia	RU	Russian Federation		
CZ	Czech Republic	LI	Liechtenstein	SD	Sudan		
DE	Germany	LK	Sri Lanka	SE	Sweden		
DK	Denmark						
EE	Estonia						

**This Page is Inserted by IFW Indexing and Scanning
Operations and is not part of the Official Record**

BEST AVAILABLE IMAGES

Defective images within this document are accurate representations of the original documents submitted by the applicant.

Defects in the images include but are not limited to the items checked:

☐ ~~BLACK BORDERS~~

☒ IMAGE CUT OFF AT TOP, BOTTOM OR SIDES

☐ FADED TEXT OR DRAWING

☒ BLURRED OR ILLEGIBLE TEXT OR DRAWING

☐ SKEWED/SLANTED IMAGES

☐ COLOR OR BLACK AND WHITE PHOTOGRAPHS

☐ GRAY SCALE DOCUMENTS

☐ LINES OR MARKS ON ORIGINAL DOCUMENT

☐ REFERENCE(S) OR EXHIBIT(S) SUBMITTED ARE POOR QUALITY

☐ OTHER: _____

IMAGES ARE BEST AVAILABLE COPY.

As rescanning these documents will not correct the image problems checked, please do not report these problems to the IFW Image Problem Mailbox.